

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100398

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/18/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.51: 0.078

LOT #: 02511080A2

EXPIRATION: 02/01/2013

BLANK TEST	0.000	12:02
INTERNAL STANDARD	VERIFIED	12:02
EXTERNAL STANDARD	0.078	12:02
BLANK TEST	0.000	12:03
EXTERNAL STANDARD	0.078	12:03
BLANK TEST	0.000	12:04
EXTERNAL STANDARD	0.078	12:05
BLANK TEST	0.000	12:05
EXTERNAL STANDARD	0.078	12:06
BLANK TEST	0.000	12:07
EXTERNAL STANDARD	0.078	12:07
BLANK TEST	0.000	12:08

Average = 0.0780

Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 1.01

PIC: 3.02

Modem: 2.0

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C

PASSED

Breath Tube = 48.1°C

PASSED

PUMP INFO

Flow Rate = 4.884 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 8/1/11


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1 day of Aug, 20 11



Notary Public's Signature

My Commission Expires on With Office



Notary Name

